

WQS Oxymeter

QUICK START

Use

- 1- Before using the device: Install the App « PreSens Wireless Studio » on a Androïd or Apple smartphone or tablet, the App is available in Google Paly or Itunes Store.
- 2- Launch the App on your smartphone and select the device that you want to connect. If no device appears, press SCAN.
- 3- Check the units (Oxygen in mg/L, temperature auto compensation) in the menu PARAMETERS
- 4- Launch the measurement in the menu LIVE VIEW
 - a- Single measurement : press Single Measurement
 - b- Continous measurement :
 - a. Define the time interval between measurement
 - b. Enable or disable data logging « Logging measurement data».
 - c. Press START
 - d. Stop measurement: only for continuous measurement press STOP.



Warning: If measurement is not stopped, the measurement will continue until the battery is discharged even you disconnect Bluetooth.

- 5- Export measurement data: select the MEASUREMENTS menu
 - a. Select the file
 - b. DOWNLOAD
 - c. EXPORT
- 6- Disconnect : press DISCONNECT in the LIVE VIEW menu.



Warning: As long as the Oxymeter is connected to your smartphone, the device is no longer visible to others smartphone or tablet.

Oxymeter not detected

If the Oxymeter is not detected by your smartphone, please check:

- 1- Bluetooth is activated on your smartphone.
- 2- The Oxymeter is nearby your smartphone (less than 10 m away).
- 3- The Oxymeter is not connected to another smartphone or tablet.
- 4- The Oxymeter still has battery power.

Battery

In case of battery replacement, press SINGLE MEASUREMENT to « refresh » the battery power level displayed.

www.vinventions.com